

## Sample Cylinders



### Stainless Steel Cylinders Pressure and Temperature Ratings

1800 psig (124 bar)  
-58 °F to 450 °F (-50 °C to 232 °C)

- DOT-3E 1800:  
75, 150, 300, and 500cc with 1/4" NPT threads
- DOT-3A 1800:  
1000 and 2250cc with 1/4" NPT threads
- DOT-3A 1800:  
3785cc (1 gallon) with 1/2" NPT threads

### Aluminum Cylinders Pressure and Temperature Rating

1800 psig (124 bar)  
0 °F to 120 °F (-18 °C to 49 °C)

- DOT-E 7737 1800:  
150, 300, and 500cc with 1/4" NPT threads

### Cylinder Design, Production and Testing

The design, manufacture and testing of steel sample cylinders is regulated by the US government in 49 CFR, Paragraphs 178.36 *Specification 3A; seamless steel cylinders* and 178.42 *Specification 3E; seamless steel cylinders*. Aluminum sample cylinders are governed by the same paragraphs, supplemented by Exemption DOT-E 7737. Specification 3A deals with cylinders not over 1,000 pounds (455 kg) water capacity and Specification 3E is for cylinders having an outside diameter no greater than 2 inches (51 mm), with a length less than 2 feet (61 cm). Service pressure is limited to 1,800 psi (124 bar) for Parker Sample Cylinders.

The above regulations control all aspects of the design and production of sample cylinders. Material physical properties and chemical characteristics are controlled. Each cylinder must be hydrostatically tested between 3,000 and 4,500 psi (207 and 310 bar). In addition, one cylinder out of each lot of 500 or less must be subjected to a burst test and result in a safety factor on burst pressure of 3.3 minimum.

All cylinder tests must be inspected and verified by an independent inspection agency, and all test reports must be maintained for fifteen years. Each cylinder must also be marked and packaged in accordance with 49 CFR.

**DOT Cylinder Retesting Requirements**  
(Per 49 CFR, Paragraph 173.34)

DOT Rating	Minimum Retest Pressure	Retest Period (Years)
3E 1800	Retest not required	N/A
3A 1800	3000 psig (207 bar)	5
E 7737 1800	Retest not required	N/A

**DOT Approved Materials for use in Aluminum Sample Cylinders**

Air, Compressed	Ethane	Methane
Ammonia, anhydrous	Ethylene	Methylamine, anhydrous
Argon	Ethylene oxide	Methyl butene
Boron trichloride	Flammable liquid, n.o.s.	Monoethylamine
Butadiene, inhibited	(Dimethylpropane)	Neon
Butane	Refrigerant gas (Freon)	Nitrogen
Carbon Dioxide, liquefied	Helium	Nitrous oxide
Compressed gas, n.o.s.	Hydrocarbon gas, non-liquefied	Oxygen*
(Bromotrifluorethylene)	(Coal gas)	Pentane
Compressed gas, n.os.	Hydrogen	Propane
(Deuterium)	Hydrogen sulphide	Sulphur dioxide
Compressed gas, n.o.s.	Isobutane	Sulphur hexafluoride
(Difluoroethylene)	Isobutylene	Trimethylamine, anhydrous
Compressed gas, n.o.s. (Krypton)	Liquefied petroleum gas (Butene)	Vinyl chloride
Compressed gas, n.o.s. (Ozone)	Liquefied petroleum gas	Vinyl fluoride, inhibited
Cyclopropane	(Cyclobutane)	Xenon
Dimethylamine, anhydrous	Liquefied petroleum gas	<b>*Note:</b> Oxygen is only acceptable if
Dimethyl ether	(Propylene)	the cylinder has straight threads

**Aluminum Sample Cylinders**

This cylinder is produced from a high strength aluminum alloy designated by the Aluminum Association as 6061 and heat treated to T6 temper. It should be noted that this cylinder is manufactured under Department of Transportation Exemption Number DOT-E 7737-1800 and is the aluminum equivalent of DOT-3E 1800 cylinder. The cylinder has been inspected by and testing has been witnessed by an independent inspection agency.

It is important that the user respect the fact that this is an aluminum cylinder. If this cylinder is exposed to fire, it should not be refilled and the cylinder should be properly disposed.

In some cases, the user may evacuate or clean the cylinder by heating in an oven; and in this case the cylinder should not be exposed to temperature above 350 °F.

This cylinder is approved for any hazardous material for which the DOT specification 3AL cylinder is prescribed or authorized in 49 CFR Part 173, classed as flammable gas, non-flammable gas, flammable liquid, or Poison A. The gases listed above are generally those that are approved.

## Oxygen Service

For aluminum sample cylinders to be charged with oxygen, the following applies: Straight threads only. (Since this cylinder has tapered threads it cannot be used for oxygen service.)

## Nitrous Oxide & Oxygen Service

Each cylinder requires special cleaning in compliance with Federal Specification RR-C-901b dated August 1, 1967, paragraph 3.8.2. This aluminum sample cylinder has not been processed by this special cleaning and therefore should not be used for oxygen service.

## Valve Insertion Procedure for Sample Cylinders

1. It is recommended that new valves be used. If a valve has been used in a steel cylinder, the threads may be distorted and possibly would not produce a proper seal.
2. Valve and cylinder threads should be clean. Examine the valve and cylinder for damaged threads, and reject or repair those containing defects such as burrs, dings, nicks, gouges, etc.
3. Apply 1-1/2 wraps of PTFE tape on the valve threads, leaving the first lead thread exposed. Apply a PTFE paste to the first lead thread and over the tape, evenly but sparingly. For stainless steel valves, PTFE paste is suggested for proper sealing.
4. Assemble the lubricated valve to the cylinder by inserting and hand tightening to engage a minimum of 2 to 3 threads. If the valve fails to start easily, recheck the valve to make sure it is to gauge. Also, check the valve and cylinder for damaged threads.
5. Place the valved cylinder in a holding device providing protective material around the cylinder to prevent gouging of the side walls with the vice holding jaws. Using a torque wrench with an adapter to fit the wrench flats on the valve, tighten the valve to **8 to 10 foot pounds maximum**. This torque should yield another 2 to 3 threads engagement, giving a total engagement of 5 to 6 threads. If a problem develops, please contact the Instrumentation Valve Division or Parker Distributor for assistance.

## Caution

Aluminum cylinders require care in the assembly of valves with tapered threads. The aluminum cylinder is of the same relative hardness as a brass valve and does not reshape or rethread the valve during insertion as does a steel cylinder. Therefore, an interface problem could be created by damaged threads or excessive valve torque preventing a gas-tight connection.


## VALVE CONNECTOR MAKE-UP INSTRUCTIONS

### MALE AND FEMALE PIPE PORTS

Wrench flats are provided on the Valve Body. It is recommended a smooth-jawed wrench or vise be used to grip the Valve Body.

1. On the male threaded part of the connection, apply a high quality pipe joint compound or PTFE tape made for this purpose. When PTFE tape is used, it is recommended two full turns of tape be applied. PTFE tape should not be overhanging or covering the first thread.
2. Engage the Valve and the other component part together, until hand-tight.
3. With a proper wrench, holding both the Valve and the component part, continue to tighten to achieve a leak-tight joint.

**F** -ANSI/ASME B1.20.1  
Internal pipe threads



**M** -ANSI/ASME B1.20.1  
External pipe threads



---

## WARNING

**FAILURE OR IMPROPER SELECTION OR IMPROPER USE OF THE PRODUCTS AND/OR SYSTEMS DESCRIBED HEREIN OR RELATED ITEMS CAN CAUSE DEATH, PERSONAL INJURY AND PROPERTY DAMAGE.**

This document and other information from Parker Hannifin Corporation, its subsidiaries and authorized distributors provide product and/or system options for further investigation by users having technical expertise. It is important that you analyze all aspects of your application and review the information concerning the product or system in the current product catalog. Due to the variety of operating conditions and applications for these products or systems, the user, through its own analysis and testing, is solely responsible for making the final selection of the products and systems and assuring that all performance, safety and warning requirements of the application are met.

The products described herein, including without limitation, product features, specifications, designs, availability and pricing, are subject to change by Parker Hannifin Corporation and its subsidiaries at any time without notice.

**ALL PARKER VALVES MUST PASS A RIGID OPERATIONAL AND LEAKAGE TEST BEFORE LEAVING THE FACTORY. IT IS RECOMMENDED AFTER ANY REASSEMBLY, THE VALVE SHOULD BE TESTED BY THE USER FOR OPERATION AND LEAKAGE. IF THESE INSTRUCTIONS ARE NOT FULLY COMPLIED WITH, THE REPAIRED PRODUCT MAY FAIL AND CAUSE DAMAGE TO PROPERTY OR INJURY TO PERSONS. PARKER HANNIFIN CANNOT ASSUME RESPONSIBILITY FOR PERFORMANCE OF A CUSTOMER SERVICED VALVE.**

